


TDSscan10 Pocket TDS Tester

# USER MANUAL




## Overview


Thank you for selecting the TDSscan series pocket TDS tester. This user manual provides a step-by-step guide to help you operate the tester, please carefully read the following instructions before use.

### Installing the Batteries


1. Remove the battery compartment lid.






2. Insert three LR44 alkaline batteries into the battery compartment, note polarity.



3. Replace the battery compartment lid to its original position and turn clockwise until tight.




### Keypad

Key	Function
	<ul style="list-style-type: none"> <li>• Switch the tester on or off</li> <li>• Lock or unlock measurement</li> <li>• Confirm or exit the calibration mode</li> <li>• Return to the TDS measurement mode</li> </ul>
	<ul style="list-style-type: none"> <li>• Press and hold the key to start the calibration</li> <li>• Increase value during the setting</li> </ul>
	<ul style="list-style-type: none"> <li>• Press and hold the key to set the TDS factor</li> <li>• Decrease value during the setting</li> <li>• Show the temperature readings</li> </ul>


## Prior to Use

Remove the protective cap from the bottom of the tester. If the sensor has dried out, soak the electrode for about 10 minutes in tap water.



## Switching the Tester On and Off

- Press the **Meas** key to switch on the tester.
- Press and hold the **Meas** key to switch off the tester.



If you do not press any key within 8 minutes, the tester will switch off automatically to conserve energy.


## TDS Calibration

The TDSscan10 tester allows 1 point calibration. For better accuracy, we recommend to calibrate the tester regularly. The following table shows acceptable standard solutions for each tester.


Model	Standard Solution Range	Default
TDSscan10L	50 to 99.9 ppm	71.8 ppm
TDSscan10M	500 to 999 ppm	744 ppm
TDSscan10H	5.00 to 9.99 ppt	7447 ppm

Make sure that using a fresh standard solution during the calibration. DO NOT reuse the standard solution after calibration, contaminants in solution will affect the calibration and eventually the accuracy of the measurement.


1. Rinse the electrode with distilled water and place into the standard solution, stir tester gently to remove air bubbles trapped in the slot of the sensor.




- Press and hold the **Cal** key to enter the calibration mode, the display shows ---.



- Press the **▲** / **▼** key to set the calibration value, make sure that the displayed value matches the calibration standard.



- Press the **Meas** key, the calibration value will automatically flash three times.



- Press the **Meas** key again, the tester returns to the measurement mode, the CAL icon disappears from the display. Calibration is completed.



- During the calibration, if the display shows **Err** indicating that the measured TDS value deviates from the theoretical value of the standard solution. The calibration will not be accepted. Please replace the fresh standard solution and calibrate the tester again.
- During the setting, press and hold the **▲** / **▼** key will make the value change faster.
- If you want to exit the calibration, DO NOT press the **Meas** key in the step 4. Press and hold the **▲** / **▼** key until the display shows ---, press the **Meas** key to exit.


## Measurement

### TDS Measurement


Rinse the electrode with distilled water, place the electrode into the sample solution and stir gently, make sure that no air bubbles on the sensor surface. Wait for the measurement to stabilize and record the reading.



- During the measurement, DO NOT completely immerse the tester in water.




- Press the **Hold** key, the tester will lock the measurement, the HOLD icon appears on the display. Press the key again to resume measurement.



- If the display shows **!** indicating the measurement exceeds the range, remove the tester from the sample immediately.

### Temperature Measurement

- Press the **°C** key, the tester shows the temperature readings.
- Press the **Meas** key to return to the TDS measurement.



## Electrode Maintenance and Replacement

- Rinse the electrode thoroughly with distilled water after use.
- Do not touch the platinum black coating on the sensor surface and always keep it clean.
- If there is a build-up of solids inside the sensor, remove carefully, then recalibrate the tester.
- If you do not use the tester for long periods, remove the batteries.

### Replacing the Electrode

If the tester fails to calibrate or gives fluctuating readings, you should consider replacing the electrode.


1. Twist the electrode collar counter clockwise, pull the electrode away from the tester.



2. Align the slot on the new electrode, gently push the electrode into the tester.



3. Twist the electrode collar clockwise until tight.



## Appendix

### Calculating the TDS Conversion Factor

The TDSscan10 tester allows setting the conversion factor according to the type of sample. The factory default is 0.50. The following formula describes the calculation method.

$$\text{Factor} = \text{Actual TDS} / \text{Actual Conductivity @ 25}^{\circ}\text{C}$$


Where:

Actual TDS = Value from the high purity water and precisely weighed NaCl or KCl reagent


Actual Conductivity = The tester measured conductivity value

### Setting the TDS Conversion Factor


1. Press and hold the °C key, the display shows ---.



2. Press the ▲/▼ key to set the conversion factor.



3. Press the Meas key to save and return to the measurement mode.



### Preparation of TDS Standard Solutions

Place the analytical grade potassium chloride (KCl) in a beaker and dry in an oven for about 3 hours at 105°C (221°F), then cool to room temperature. Add the reagent to a 1 liter volumetric flask according to the instructions in table below.

TDS Standard	Reagent	Weight
71.8 ppm	KCl	74.4 mg
744.7 ppm	KCl	745.5 mg
7447 ppm	KCl	7.45 g

Fill the distilled water to the mark, mix the solution until the reagent is completely dissolved.

### Optional Accessories

Order Code	Description
E-ECscan-C1-10K	2-pole conductivity cell, K=1
TDSCS-718	Standard solution 71.8 ppm, 480 ml
TDSCS-744	Standard solution 744 ppm, 480 ml
TDSCS-7447	Standard solution 7447 ppm, 480 ml

## Tester Specifications

Model	TDSscan10
<b>TDS</b>	
Range	TDSscan10L: 0.5 to 100.0 ppm
	TDSscan10M: 5 to 1000 ppm
	TDSscan10H: 0.05 to 10.00 ppt
Resolution	TDSscan10L: 0.1 ppm
	TDSscan10M: 1 ppm
	TDSscan10H: 0.01 ppt
Accuracy	±1% F.S
Calibration Point	1 point
Calibration Solution	TDSscan10L: 71.8 ppm
	TDSscan10M: 744.7 ppm
	TDSscan10H: 7.447 ppt
Temperature Compensation	0 to 50°C, automatic
TDS Factor	0.4 to 1.0
<b>Temperature</b>	
Range	0 to 50°C
Resolution	1°C
Accuracy	±1°C
<b>Other Specifications</b>	
Operating Temperature	0 to 50°C (32 to 122°F)
Storage Temperature	0 to 60°C (32 to 140°F)
Relative Humidity	< 80% (non-condensing)
IP Rating	IP65
Display	LCD, 21 × 21 mm (0.82 × 0.82 in.)
Power Requirements	3 × 1.5V LR44 micro alkaline batteries
Auto-Off	8 minutes after last key pressed
Dimensions	185 (L) × 40 (Ø) mm (7.28 × 1.57 in.)
Weight	100g (3.5 oz.)

## Disposal

This product is required to comply with the European Union's Waste Electrical and Electronic Equipment (WEEE) Directive 2002/96/EC and may not be disposed of in domestic waste. Please dispose of product in accordance with local regulations at the collecting point specified for electrical and electronic equipment.



## Warranty

The warranty period for tester is one year from the date of shipment. Above warranty does not cover the electrode and standard solution.

Out of warranty products will be repaired on a charged basis.

The warranty on your tester shall not apply to defects resulting from:

- Improper or inadequate maintenance by customer
- Unauthorized modification or misuse
- Operation outside of the environment specifications of the products

For more information, please contact the supplier.



Office: 4715 Castlewood St., Sugar land, TX 77479, USA

Tel: (+1) 346-762-7358

E-mail: banteinstruments@yahoo.com

Factory: 2185 Laifang Rd., Shanghai 201615, China

Tel: (+86) 21-6404-1598

E-mail: banteinstrument@hotmail.com

 [www.bante-china.com](http://www.bante-china.com)



The information in this document is subject to change without notice.  
Copyright © Bante Instruments Inc, 2022. All rights reserved.